

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
 Data File : VY005579.D
 Acq On : 06 Aug 2021 18:30
 Operator : SY/MD
 Sample : M3295-02
 Misc : 3.91G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FK-EF-01-022-ABCD

Quant Time: Aug 07 06:21:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	5881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	9300	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	8527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2574	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	3951	66.37	ug/l	0.00
Spiked Amount 50.000	Range	50 - 163	Recovery	= 132.74%		
35) Dibromofluoromethane	7.716	113	2037	39.12	ug/l	-0.01
Spiked Amount 50.000	Range	54 - 147	Recovery	= 78.24%		
50) Toluene-d8	10.179	98	11526	52.78	ug/l	0.00
Spiked Amount 50.000	Range	49 - 140	Recovery	= 105.56%		
62) 4-Bromofluorobenzene	12.483	95	3915	54.14	ug/l	0.00
Spiked Amount 50.000	Range	25 - 144	Recovery	= 108.28%		
Target Compounds						Qvalue
10) Methyl Iodide	4.107	142	549	8.78	ug/l #	86
20) Methylene Chloride	4.729	84	999	12.21	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
Data File : VY005579.D
Acq On : 06 Aug 2021 18:30
Operator : SY/MD
Sample : M3295-02
Misc : 3.91G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FK-EF-01-022-ABCD

Quant Time: Aug 07 06:21:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

