

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080723\
 Data File : VY014970.D
 Acq On : 07 Aug 2023 18:02
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Aug 08 07:53:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 07:44:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	216068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	377649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	370351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	181408	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	139429	53.241	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.480%
35) Dibromofluoromethane	7.716	113	119304	49.402	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.179	98	452187	49.194	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.477	95	173510	53.260	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080723\
Data File : VY014970.D
Acq On : 07 Aug 2023 18:02
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Aug 08 07:53:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 07:44:56 2023
Response via : Initial Calibration

