

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080823\
 Data File : VY014974.D
 Acq On : 08 Aug 2023 10:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 15:51:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 07:44:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.305	0.254	16.7	93	0.00
3 P	Chloromethane	0.393	0.353	10.2	95	0.00
4 C	Vinyl Chloride	0.485	0.431	11.1#	93	0.00
5 T	Bromomethane	0.368	0.342	7.1	102	0.00
6 T	Chloroethane	0.335	0.318	5.1	96	0.00
7 T	Trichlorofluoromethane	0.766	0.687	10.3	91	0.00
8 T	Diethyl Ether	0.283	0.279	1.4	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.454	0.415	8.6	91	0.01
10 T	Methyl Iodide	0.469	0.485	-3.4	96	0.00
11 T	Tert butyl alcohol	0.146	0.113	22.6	113	0.01
12 CM	1,1-Dichloroethene	0.395	0.374	5.3#	95	0.00
13 T	Acrolein	0.050	0.061	-22.0	131	0.00
14 T	Allyl chloride	0.658	0.649	1.4	97	0.00
15 T	Acrylonitrile	0.173	0.174	-0.6	99	0.00
16 T	Acetone	0.189	0.214	-13.2	110	0.00
17 T	Carbon Disulfide	0.874	0.803	8.1	96	0.00
18 T	Methyl Acetate	0.688	0.706	-2.6	102	0.01
19 T	Methyl tert-butyl Ether	1.495	1.501	-0.4	98	0.01
20 T	Methylene Chloride	1.162	0.704	39.4#	96	0.00
21 T	trans-1,2-Dichloroethene	0.450	0.435	3.3	97	0.00
22 T	Diisopropyl ether	1.533	1.549	-1.0	98	0.00
23 T	Vinyl Acetate	0.922	0.930	-0.9	97	0.00
24 P	1,1-Dichloroethane	0.870	0.866	0.5	98	0.00
25 T	2-Butanone	0.271	0.277	-2.2	102	0.00
26 T	2,2-Dichloropropane	0.880	0.813	7.6	94	0.01
27 T	cis-1,2-Dichloroethene	0.570	0.566	0.7	98	0.00
28 T	Bromochloromethane	0.392	0.377	3.8	93	0.00
29 T	Tetrahydrofuran	0.159	0.160	-0.6	101	0.00
30 C	Chloroform	0.945	0.953	-0.8#	97	0.00
31 T	Cyclohexane	0.706	0.604	14.4	93	0.00
32 T	1,1,1-Trichloroethane	0.832	0.812	2.4	95	0.00
33 S	1,2-Dichloroethane-d4	0.606	0.810	-33.7#	130	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.320	0.435	-35.9#	132	0.00
36 T	1,1-Dichloropropene	0.397	0.372	6.3	94	0.00
37 T	Ethyl Acetate	0.315	0.308	2.2	100	0.00
38 T	Carbon Tetrachloride	0.446	0.422	5.4	93	0.00
39 T	Methylcyclohexane	0.472	0.428	9.3	92	0.00
40 TM	Benzene	1.169	1.124	3.8	97	0.00
41 T	Methacrylonitrile	0.153	0.185	-20.9	119	0.02
42 TM	1,2-Dichloroethane	0.407	0.402	1.2	98	0.00
43 T	Isopropyl Acetate	0.585	0.578	1.2	99	0.00
44 TM	Trichloroethene	0.335	0.321	4.2	96	0.00
45 C	1,2-Dichloropropane	0.311	0.305	1.9#	97	0.00
46 T	Dibromomethane	0.200	0.198	1.0	98	0.00
47 T	Bromodichloromethane	0.461	0.465	-0.9	98	0.00
48 T	Methyl methacrylate	0.257	0.257	0.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080823\
 Data File : VY014974.D
 Acq On : 08 Aug 2023 10:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 15:51:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 07:44:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	100	0.00
50 S	Toluene-d8	1.217	1.592	-30.8#	129	0.00
51 T	4-Methyl-2-Pentanone	0.330	0.330	0.0	99	0.00
52 CM	Toluene	0.752	0.729	3.1#	97	0.00
53 T	t-1,3-Dichloropropene	0.480	0.491	-2.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.519	0.518	0.2	96	0.00
55 T	1,1,2-Trichloroethane	0.291	0.289	0.7	99	0.00
56 T	Ethyl methacrylate	0.416	0.423	-1.7	98	0.00
57 T	1,3-Dichloropropane	0.481	0.475	1.2	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.197	0.199	-1.0	98	0.00
59 T	2-Hexanone	0.244	0.251	-2.9	102	0.00
60 T	Dibromochloromethane	0.346	0.348	-0.6	97	0.00
61 T	1,2-Dibromoethane	0.278	0.275	1.1	99	0.00
62 S	4-Bromofluorobenzene	0.431	0.574	-33.2#	132	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.376	0.360	4.3	94	0.00
65 PM	Chlorobenzene	0.949	0.916	3.5	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.368	1.1	97	0.00
67 C	Ethyl Benzene	1.665	1.593	4.3#	96	0.00
68 T	m/p-Xylenes	0.637	0.608	4.6	96	0.00
69 T	o-Xylene	0.623	0.604	3.0	97	0.00
70 T	Styrene	1.065	1.045	1.9	97	0.00
71 P	Bromoform	0.256	0.257	-0.4	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.361	3.183	5.3	96	0.00
74 T	N-amyl acetate	1.148	1.132	1.4	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.821	0.794	3.3	99	0.00
76 T	1,2,3-Trichloropropane	0.599	0.571	4.7	99	0.00
77 T	Bromobenzene	0.815	0.779	4.4	98	0.00
78 T	n-propylbenzene	4.020	3.756	6.6	95	0.00
79 T	2-Chlorotoluene	2.291	2.180	4.8	97	0.00
80 T	1,3,5-Trimethylbenzene	2.770	2.641	4.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.272	0.274	-0.7	100	0.00
82 T	4-Chlorotoluene	2.368	2.268	4.2	97	0.00
83 T	tert-Butylbenzene	2.534	2.400	5.3	96	0.00
84 T	1,2,4-Trimethylbenzene	2.775	2.641	4.8	96	0.00
85 T	sec-Butylbenzene	3.719	3.488	6.2	95	0.00
86 T	p-Isopropyltoluene	3.078	2.939	4.5	96	0.00
87 T	1,3-Dichlorobenzene	1.596	1.535	3.8	98	0.00
88 T	1,4-Dichlorobenzene	1.591	1.531	3.8	98	0.00
89 T	n-Butylbenzene	2.892	2.754	4.8	95	0.00
90 T	Hexachloroethane	0.598	0.572	4.3	96	0.00
91 T	1,2-Dichlorobenzene	1.473	1.437	2.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.160	0.157	1.9	101	0.00
93 T	1,2,4-Trichlorobenzene	0.942	0.946	-0.4	99	0.00
94 T	Hexachlorobutadiene	0.511	0.491	3.9	95	0.00
95 T	Naphthalene	2.218	2.259	-1.8	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080823\
Data File : VY014974.D
Acq On : 08 Aug 2023 10:25
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 08 15:51:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 07:44:56 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.853	0.854	-0.1	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6