

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
 Data File : VY005585.D
 Acq On : 10 Aug 2021 11:09
 Operator : SY/MD
 Sample : M3341-01
 Misc : 5.04G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 228-344

Quant Time: Aug 11 03:36:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	198814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	318502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	300741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	137597	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	106584	52.961	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.920%
35) Dibromofluoromethane	7.728	113	92322	51.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.540%
50) Toluene-d8	10.185	98	400556	53.560	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.120%
62) 4-Bromofluorobenzene	12.483	95	128417	51.853	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
Data File : VY005585.D
Acq On : 10 Aug 2021 11:09
Operator : SY/MD
Sample : M3341-01
Misc : 5.04G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
228-344

Quant Time: Aug 11 03:36:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

