

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
 Data File : VY005586.D
 Acq On : 10 Aug 2021 11:32
 Operator : SY/MD
 Sample : M3340-01
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 205-239-929

Quant Time: Aug 11 03:36:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	235296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	392529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	385599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	196783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	134247	56.364	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.720%
35) Dibromofluoromethane	7.728	113	117585	53.504	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.000%
50) Toluene-d8	10.185	98	495968	53.811	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.620%
62) 4-Bromofluorobenzene	12.483	95	171249	56.107	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.220%
Target Compounds						
20) Methylene Chloride	4.728	84	34603	9.643	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
Data File : VY005586.D
Acq On : 10 Aug 2021 11:32
Operator : SY/MD
Sample : M3340-01
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
205-239-929

Quant Time: Aug 11 03:36:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

