

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
 Data File : VY005591.D
 Acq On : 10 Aug 2021 13:30
 Operator : SY/MD
 Sample : M3314-01RE
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11931RE

Quant Time: Aug 11 03:36:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	211660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	355584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	350885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	174176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	126109	58.860	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 117.720%	
35) Dibromofluoromethane	7.728	113	107159	53.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.660%	
50) Toluene-d8	10.185	98	450669	53.976	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	12.483	95	155055	56.080	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.160%	
Target Compounds						
20) Methylene Chloride	4.729	84	32360	10.298	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
Data File : VY005591.D
Acq On : 10 Aug 2021 13:30
Operator : SY/MD
Sample : M3314-01RE
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11931RE

Quant Time: Aug 11 03:36:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

