

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
 Data File : VY005597.D
 Acq On : 10 Aug 2021 15:49
 Operator : SY/MD
 Sample : M3331-02
 Misc : 6.33G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BSY1-SB-10-3-4

Quant Time: Aug 11 03:37:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	217701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	364092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	372748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	188572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	133494	60.578	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.160%
35) Dibromofluoromethane	7.722	113	108975	53.459	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.920%
50) Toluene-d8	10.185	98	471425	55.143	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.280%
62) 4-Bromofluorobenzene	12.483	95	168811	59.628	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.260%
Target Compounds						Qvalue
16) Acetone	3.960	43	5614	10.104	ug/l	98
95) Naphthalene	15.239	128	8478	1.123	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
Data File : VY005597.D
Acq On : 10 Aug 2021 15:49
Operator : SY/MD
Sample : M3331-02
Misc : 6.33G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BSY1-SB-10-3-4

Quant Time: Aug 11 03:37:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

