

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
 Data File : VY005598.D
 Acq On : 10 Aug 2021 16:12
 Operator : SY/MD
 Sample : M3333-01
 Misc : 4.35G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 124055

Quant Time: Aug 11 03:37:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	158948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	261404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	254076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129169	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	90747	56.401	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.800%
35) Dibromofluoromethane	7.722	113	76904	52.547	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.100%
50) Toluene-d8	10.185	98	330518	53.848	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.483	95	114256	56.212	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081021\
Data File : VY005598.D
Acq On : 10 Aug 2021 16:12
Operator : SY/MD
Sample : M3333-01
Misc : 4.35G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
124055

Quant Time: Aug 11 03:37:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

