



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081022\
Data File : VY009949.D
Acq On : 10 Aug 2022 09:02
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
Title : SW846 8260
Last Update : Wed Aug 10 17:37:28 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	2677	PASS
75	95	30	60	50.0	7235	PASS
95	95	100	100	100.0	14459	PASS
96	95	5	9	7.2	1042	PASS
173	174	0.00	2	1.3	158	PASS
174	95	50	100	86.7	12534	PASS
175	174	5	9	7.5	941	PASS
176	174	95	101	95.4	11957	PASS
177	176	5	9	6.8	817	PASS