

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081023\
 Data File : VY015041.D
 Acq On : 10 Aug 2023 20:34
 Operator : SY/MD
 Sample : 03955-13
 Misc : 3.82g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-26

Quant Time: Aug 11 01:58:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 07:44:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	219970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	374147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	368591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	189448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	137446	51.553	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.100%
35) Dibromofluoromethane	7.716	113	118314	49.450	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.900%
50) Toluene-d8	10.179	98	440118	48.329	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.660%
62) 4-Bromofluorobenzene	12.483	95	174610	54.099	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081023\
Data File : VY015041.D
Acq On : 10 Aug 2023 20:34
Operator : SY/MD
Sample : 03955-13
Misc : 3.82g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-26

Quant Time: Aug 11 01:58:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 07:44:56 2023
Response via : Initial Calibration

