



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081023\
 Data File : VY015018.D
 Acq On : 10 Aug 2023 10:42
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080723S.M
 Title : SW846 8260
 Last Update : Tue Aug 08 07:44:56 2023



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	37141	PASS
75	95	30	60	53.4	100483	PASS
95	95	100	100	100.0	188309	PASS
96	95	5	9	7.0	13176	PASS
173	174	0.00	2	1.1	1628	PASS
174	95	50	100	81.4	153365	PASS
175	174	5	9	7.8	12038	PASS
176	174	95	101	95.9	147003	PASS
177	176	5	9	6.9	10150	PASS