

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081221\
 Data File : VY005614.D
 Acq On : 12 Aug 2021 12:00
 Operator : SY/MD
 Sample : M3345-02
 Misc : 5.29G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DRUM-3-18B

Quant Time: Aug 13 04:20:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	147325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	235546	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	223792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	109426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	77739	52.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.260%	
35) Dibromofluoromethane	7.728	113	67129	50.903	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.800%	
50) Toluene-d8	10.185	98	302097	54.621	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.240%	
62) 4-Bromofluorobenzene	12.483	95	99361	54.250	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.500%	
Target Compounds						Qvalue
16) Acetone	3.954	43	5037	13.397	ug/l	92
17) Carbon Disulfide	4.198	76	15312	3.912	ug/l #	93
39) Methylcyclohexane	9.185	83	6762	2.406	ug/l	86
84) 1,2,4-Trimethylbenzene	13.117	105	9010	1.440	ug/l	96
86) p-Isopropyltoluene	13.373	119	16940	2.405	ug/l	93
95) Naphthalene	15.239	128	14911	3.405	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081221\
Data File : VY005614.D
Acq On : 12 Aug 2021 12:00
Operator : SY/MD
Sample : M3345-02
Misc : 5.29G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DRUM-3-18B

Quant Time: Aug 13 04:20:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

