

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081221\
 Data File : VY005626.D
 Acq On : 12 Aug 2021 16:38
 Operator : SY/MD
 Sample : M3343-08
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20210582-COM-5

Quant Time: Aug 13 04:23:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	243425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	417706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	426539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	223555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	161735	65.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 131.280%	
35) Dibromofluoromethane	7.728	113	127527	54.531	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.060%	
50) Toluene-d8	10.185	98	537109	54.762	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.520%	
62) 4-Bromofluorobenzene	12.483	95	201462	62.028	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 124.060%	
Target Compounds						Qvalue
16) Acetone	3.954	43	63029	101.455	ug/l	97
17) Carbon Disulfide	4.204	76	42912	6.636	ug/l	99
25) 2-Butanone	6.996	43	20984	22.850	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081221\
Data File : VY005626.D
Acq On : 12 Aug 2021 16:38
Operator : SY/MD
Sample : M3343-08
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20210582-COM-5

Quant Time: Aug 13 04:23:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

