

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
 Data File : VY005638.D
 Acq On : 13 Aug 2021 16:52
 Operator : SY/MD
 Sample : M3408-01
 Misc : 9.6G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 31S

Quant Time: Aug 14 03:36:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	230817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	391527	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	392971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	201976	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	148179	63.421	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.840%
35) Dibromofluoromethane	7.728	113	117078	53.410	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.820%
50) Toluene-d8	10.185	98	499269	54.307	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.620%
62) 4-Bromofluorobenzene	12.483	95	180494	59.288	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
Data File : VY005638.D
Acq On : 13 Aug 2021 16:52
Operator : SY/MD
Sample : M3408-01
Misc : 9.6G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
31S

Quant Time: Aug 14 03:36:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

