

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
 Data File : VY005641.D
 Acq On : 13 Aug 2021 18:02
 Operator : SY/MD
 Sample : M3408-10
 Misc : 9.24G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 34S

Quant Time: Aug 14 03:37:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	217706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	369712	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	377699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	195020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	140094	63.571	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.140%	
35) Dibromofluoromethane	7.728	113	114377	55.257	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.520%	
50) Toluene-d8	10.185	98	482728	55.607	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.220%	
62) 4-Bromofluorobenzene	12.483	95	172642	60.054	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.100%	
Target Compounds						Qvalue
16) Acetone	3.954	43	7475	13.454	ug/l	# 84
18) Methyl Acetate	4.485	43	1666	1.146	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
Data File : VY005641.D
Acq On : 13 Aug 2021 18:02
Operator : SY/MD
Sample : M3408-10
Misc : 9.24G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
34S

Quant Time: Aug 14 03:37:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

