

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
 Data File : VY005645.D
 Acq On : 13 Aug 2021 19:34
 Operator : SY/MD
 Sample : M3368-04
 Misc : 5.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-01-1015

Quant Time: Aug 14 03:38:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	210641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	358971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	366973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	188282	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	130627	61.264	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.520%
35) Dibromofluoromethane	7.728	113	108646	54.058	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.120%
50) Toluene-d8	10.185	98	461043	54.698	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.400%
62) 4-Bromofluorobenzene	12.483	95	164288	58.858	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081321\
Data File : VY005645.D
Acq On : 13 Aug 2021 19:34
Operator : SY/MD
Sample : M3368-04
Misc : 5.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-01-1015

Quant Time: Aug 14 03:38:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

