

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081325\
 Data File : VY023076.D
 Acq On : 13 Aug 2025 17:40
 Operator : SY/MD
 Sample : Q2819-14
 Misc : 6.72g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 17M-W

Quant Time: Aug 14 01:27:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	675805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	1293758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1059306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	379138	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	327850	50.901	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.800%
35) Dibromofluoromethane	7.628	113	390959	50.502	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.000%
50) Toluene-d8	10.103	98	1562265	51.658	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.320%
62) 4-Bromofluorobenzene	12.401	95	401739	41.306	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081325\
Data File : VY023076.D
Acq On : 13 Aug 2025 17:40
Operator : SY/MD
Sample : Q2819-14
Misc : 6.72g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
17M-W

Quant Time: Aug 14 01:27:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

