

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
 Data File : VY010008.D
 Acq On : 14 Aug 2022 19:27
 Operator : KP/MD
 Sample : N4189-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB20

Quant Time: Aug 15 00:41:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	11887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	20548	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	17711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	6500	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5162	42.809	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.620%
35) Dibromofluoromethane	7.710	113	5443	43.280	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.560%
50) Toluene-d8	10.179	98	19365	40.187	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	80.380%
62) 4-Bromofluorobenzene	12.477	95	5480	31.039	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
Data File : VY010008.D
Acq On : 14 Aug 2022 19:27
Operator : KP/MD
Sample : N4189-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB20

Quant Time: Aug 15 00:41:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 13 02:54:57 2022
Response via : Initial Calibration

