

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
 Data File : VY010020.D
 Acq On : 15 Aug 2022 00:08
 Operator : KP/MD
 Sample : N4046-09
 Misc : 6.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BEACONPOINT-BACKFILL-05

Quant Time: Aug 15 00:43:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	72393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	137882	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	139445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	39703	54.065	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.120%
35) Dibromofluoromethane	7.710	113	38672	45.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.660%
50) Toluene-d8	10.179	98	138862	42.945	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.880%
62) 4-Bromofluorobenzene	12.477	95	48237	40.716	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
Data File : VY010020.D
Acq On : 15 Aug 2022 00:08
Operator : KP/MD
Sample : N4046-09
Misc : 6.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BEACONPOINT-BACKFILL-05

Quant Time: Aug 15 00:43:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 13 02:54:57 2022
Response via : Initial Calibration

