

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
 Data File : VY010022.D
 Acq On : 15 Aug 2022 00:54
 Operator : KP/MD
 Sample : N4046-13
 Misc : 7.49g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BEACONPOINT-BACKFILL-07

Quant Time: Aug 15 06:24:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	73553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	141780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	142196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	65190	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40432	54.189	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.380%
35) Dibromofluoromethane	7.710	113	40384	46.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.080%
50) Toluene-d8	10.179	98	145138	43.652	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.300%
62) 4-Bromofluorobenzene	12.477	95	49156	40.351	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
Data File : VY010022.D
Acq On : 15 Aug 2022 00:54
Operator : KP/MD
Sample : N4046-13
Misc : 7.49g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BEACONPOINT-BACKFILL-07

Quant Time: Aug 15 06:24:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 13 02:54:57 2022
Response via : Initial Calibration

