



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081422\
 Data File : VY010001.D
 Acq On : 14 Aug 2022 15:48
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Title : SW846 8260
 Last Update : Sat Aug 13 02:54:57 2022



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	7767	PASS
75	95	30	60	46.4	23091	PASS
95	95	100	100	100.0	49803	PASS
96	95	5	9	6.8	3397	PASS
173	174	0.00	2	1.5	555	PASS
174	95	50	100	76.6	38173	PASS
175	174	5	9	8.0	3063	PASS
176	174	95	101	95.4	36419	PASS
177	176	5	9	6.5	2374	PASS