

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081423\
 Data File : VY015078.D
 Acq On : 14 Aug 2023 14:49
 Operator : SY/MD
 Sample : 03974-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-081023

Quant Time: Aug 14 15:02:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	18504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	30086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	25677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	12178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	10227	46.042	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.080%		
35) Dibromofluoromethane	7.716	113	8531	44.076	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	88.160%		
50) Toluene-d8	10.173	98	33930	47.091	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	94.180%		
62) 4-Bromofluorobenzene	12.477	95	11271	42.828	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	85.660%		
Target Compounds						
20) Methylene Chloride	4.698	84	1384	3.978	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081423\
Data File : VY015078.D
Acq On : 14 Aug 2023 14:49
Operator : SY/MD
Sample : 03974-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-081023

Quant Time: Aug 14 15:02:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

