

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081423\
 Data File : VY015079.D
 Acq On : 14 Aug 2023 15:12
 Operator : SY/MD
 Sample : 03971-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-PP02A

Quant Time: Aug 14 15:41:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	152331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	264129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	254184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	126960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	97208	53.160	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.320%	
35) Dibromofluoromethane	7.716	113	83616	49.208	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.420%	
50) Toluene-d8	10.179	98	314556	49.728	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.460%	
62) 4-Bromofluorobenzene	12.477	95	117367	50.800	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.600%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	21913	12.017	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081423\
Data File : VY015079.D
Acq On : 14 Aug 2023 15:12
Operator : SY/MD
Sample : 03971-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-PP02A

Quant Time: Aug 14 15:41:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

