

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
 Data File : VY015107.D
 Acq On : 15 Aug 2023 16:42
 Operator : SY/MD
 Sample : 04003-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-081423

Quant Time: Aug 16 02:06:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	9936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	15407	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	14088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	6578	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5283	44.294	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.580%
35) Dibromofluoromethane	7.710	113	3871	39.054	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.100%
50) Toluene-d8	10.179	98	17562	47.597	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.200%
62) 4-Bromofluorobenzene	12.477	95	6234	46.257	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
Data File : VY015107.D
Acq On : 15 Aug 2023 16:42
Operator : SY/MD
Sample : 04003-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-081423

Quant Time: Aug 16 02:06:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

