

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
 Data File : VY015120.D
 Acq On : 15 Aug 2023 22:34
 Operator : SY/MD
 Sample : VY0815SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0815SBL02

Quant Time: Aug 16 01:34:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	141659	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.691	114	245704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	238472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	90064	52.964	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.920%
35) Dibromofluoromethane	7.722	113	78234	49.493	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.980%
50) Toluene-d8	10.179	98	293744	49.920	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.483	95	111641	51.945	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.900%
Target Compounds						Qvalue
16) Acetone	3.948	43	21325	38.387	ug/l	100
20) Methylene Chloride	4.716	84	14059	6.825	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
Data File : VY015120.D
Acq On : 15 Aug 2023 22:34
Operator : SY/MD
Sample : VY0815SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0815SBL02

Quant Time: Aug 16 01:34:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

