

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
 Data File : VY015124.D
 Acq On : 16 Aug 2023 00:08
 Operator : SY/MD
 Sample : 03956-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-PP23A

Quant Time: Aug 16 04:25:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	17586	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.691	114	27258	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	22377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7487	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	7778	36.844	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	73.680%
35) Dibromofluoromethane	7.722	113	7065	40.289	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.580%
50) Toluene-d8	10.185	98	31083	47.616	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.240%
62) 4-Bromofluorobenzene	12.483	95	7790	32.672	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
Data File : VY015124.D
Acq On : 16 Aug 2023 00:08
Operator : SY/MD
Sample : 03956-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-PP23A

Quant Time: Aug 16 04:25:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

