

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
 Data File : VY015130.D
 Acq On : 16 Aug 2023 02:29
 Operator : SY/MD
 Sample : 03973-07
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-11.5-12.0

Quant Time: Aug 16 04:26:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	13852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	22078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	20529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	9431	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	8304	49.940	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.880%
35) Dibromofluoromethane	7.722	113	6374	44.876	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.760%
50) Toluene-d8	10.185	98	25047	47.371	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.483	95	8929	46.236	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081523\
Data File : VY015130.D
Acq On : 16 Aug 2023 02:29
Operator : SY/MD
Sample : 03973-07
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-11.5-12.0

Quant Time: Aug 16 04:26:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

