

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005662.D
 Acq On : 16 Aug 2021 12:03
 Operator : SY/MD
 Sample : M3404-04
 Misc : 5.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP2-1-6FT

Quant Time: Aug 17 05:17:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	177196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	287535	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	290514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	105539	58.840	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.680%
35) Dibromofluoromethane	7.728	113	87164	54.145	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.280%
50) Toluene-d8	10.185	98	371248	54.987	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.980%
62) 4-Bromofluorobenzene	12.483	95	125624	56.188	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.380%
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	21722	6.889	ug/l	90
44) Trichloroethene	8.947	130	2498	1.185	ug/l	80
64) Tetrachloroethene	10.721	164	5421	2.451	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005662.D
Acq On : 16 Aug 2021 12:03
Operator : SY/MD
Sample : M3404-04
Misc : 5.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP2-1-6FT

Quant Time: Aug 17 05:17:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

