

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005663.D
 Acq On : 16 Aug 2021 12:26
 Operator : SY/MD
 Sample : M3404-07RE
 Misc : 5.10G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP3-1-6FTRE

Quant Time: Aug 17 05:18:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	16921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	26331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	25569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10464	61.092	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.180%	
35) Dibromofluoromethane	7.722	113	7981	54.137	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.280%	
50) Toluene-d8	10.185	98	33057	53.467	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.940%	
62) 4-Bromofluorobenzene	12.483	95	10247	50.049	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.100%	
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	2007	6.442	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005663.D
Acq On : 16 Aug 2021 12:26
Operator : SY/MD
Sample : M3404-07RE
Misc : 5.10G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP3-1-6FTRE

Quant Time: Aug 17 05:18:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

