

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005667.D
 Acq On : 16 Aug 2021 13:59
 Operator : SY/MD
 Sample : M3407-10RE
 Misc : 7.15G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 26SRE

Quant Time: Aug 17 05:18:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	18003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	20096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	9262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	2949	16.182	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	32.360%#
35) Dibromofluoromethane	7.728	113	3242	28.815	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	57.620%
50) Toluene-d8	10.179	98	21490	45.542	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.080%
62) 4-Bromofluorobenzene	12.483	95	2563	16.402	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	32.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005667.D
Acq On : 16 Aug 2021 13:59
Operator : SY/MD
Sample : M3407-10RE
Misc : 7.15G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
26SRE

Quant Time: Aug 17 05:18:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

