

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005668.D
 Acq On : 16 Aug 2021 14:22
 Operator : SY/MD
 Sample : M3407-13
 Misc : 8.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 27N

Quant Time: Aug 17 05:18:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	169481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	282393	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	253660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95605	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	116350	67.820	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.640%
35) Dibromofluoromethane	7.728	113	88040	55.685	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.360%
50) Toluene-d8	10.185	98	345795	52.150	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.300%
62) 4-Bromofluorobenzene	12.483	95	101731	46.330	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005668.D
Acq On : 16 Aug 2021 14:22
Operator : SY/MD
Sample : M3407-13
Misc : 8.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
27N

Quant Time: Aug 17 05:18:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

