

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005680.D
 Acq On : 16 Aug 2021 19:00
 Operator : SY/MD
 Sample : M3429-01
 Misc : 9.92G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Aug 17 05:21:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	207507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	361191	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	322289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	162268	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	134356	63.964	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.920%
35) Dibromofluoromethane	7.722	113	109032	53.917	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.840%
50) Toluene-d8	10.185	98	407551	48.054	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.100%
62) 4-Bromofluorobenzene	12.483	95	138806	49.424	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005680.D
Acq On : 16 Aug 2021 19:00
Operator : SY/MD
Sample : M3429-01
Misc : 9.92G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Aug 17 05:21:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

