

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005695.D
 Acq On : 17 Aug 2021 02:21
 Operator : SY/MD
 Sample : M3425-01
 Misc : 8.79G/5ML/MSVOA_Y/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20S

Quant Time: Aug 17 05:24:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	179137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	310998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	324491	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169979	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	124898	68.878	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.760%	
35) Dibromofluoromethane	7.722	113	95753	54.992	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.980%	
50) Toluene-d8	10.185	98	402536	55.123	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.240%	
62) 4-Bromofluorobenzene	12.483	95	146144	60.435	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.860%	
Target Compounds						
16) Acetone	3.960	43	5683	12.431	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005695.D
Acq On : 17 Aug 2021 02:21
Operator : SY/MD
Sample : M3425-01
Misc : 8.79G/5ML/MSVOA_Y/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20S

Quant Time: Aug 17 05:24:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

