

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005686.D
 Acq On : 16 Aug 2021 22:07
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 05:22:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.411	0.414	-0.7	73	0.00
3 P	Chloromethane	0.491	0.521	-6.1	71	0.00
4 C	Vinyl Chloride	0.588	0.659	-12.1#	74	0.00
5 T	Bromomethane	0.387	0.435	-12.4	76	0.00
6 T	Chloroethane	0.355	0.434	-22.3	76	0.00
7 T	Trichlorofluoromethane	0.787	0.892	-13.3	72	0.00
8 T	Diethyl Ether	0.275	0.318	-15.6	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.427	0.499	-16.9	79	0.00
10 T	Methyl Iodide	0.532	0.632	-18.8	79	0.00
11 T	Tert butyl alcohol	0.050	0.052	-4.0	85	0.00
12 CM	1,1-Dichloroethene	0.431	0.506	-17.4#	79	-0.01
13 T	Acrolein	0.030	0.030	0.0	67	0.00
14 T	Allyl chloride	0.755	0.931	-23.3	83	0.00
15 T	Acrylonitrile	0.141	0.164	-16.3	87	0.00
16 T	Acetone	0.128	0.126	1.6	69	0.00
17 T	Carbon Disulfide	1.328	1.471	-10.8	76	0.00
18 T	Methyl Acetate	0.334	0.383	-14.7	91	0.00
19 T	Methyl tert-butyl Ether	1.335	1.601	-19.9	84	0.00
20 T	Methylene Chloride	0.641	0.675	-5.3	87	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.573	-18.9	81	0.00
22 T	Diisopropyl ether	1.557	1.801	-15.7	78	0.00
23 T	Vinyl Acetate	1.036	1.234	-19.1	83	0.00
24 P	1,1-Dichloroethane	0.863	1.093	-26.7#	85	0.00
25 T	2-Butanone	0.189	0.185	2.1	72	0.00
26 T	2,2-Dichloropropane	0.788	0.784	0.5	66	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.588	-7.3	72	0.00
28 T	Bromochloromethane	0.344	0.357	-3.8	77	0.00
29 T	Tetrahydrofuran	0.125	0.125	0.0	77	0.00
30 C	Chloroform	0.881	0.980	-11.2#	75	0.00
31 T	Cyclohexane	0.868	0.830	4.4	67	0.00
32 T	1,1,1-Trichloroethane	0.810	0.885	-9.3	73	0.00
33 S	1,2-Dichloroethane-d4	0.506	0.558	-10.3	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	68	0.00
35 S	Dibromofluoromethane	0.280	0.303	-8.2	70	0.00
36 T	1,1-Dichloropropene	0.469	0.482	-2.8	71	0.00
37 T	Ethyl Acetate	0.284	0.280	1.4	76	0.00
38 T	Carbon Tetrachloride	0.510	0.537	-5.3	72	0.00
39 T	Methylcyclohexane	0.597	0.579	3.0	66	0.00
40 TM	Benzene	1.343	1.417	-5.5	73	0.00
41 T	Methacrylonitrile	0.154	0.172	-11.7	72	0.00
42 TM	1,2-Dichloroethane	0.436	0.472	-8.3	78	0.00
43 T	Isopropyl Acetate	0.538	0.547	-1.7	77	0.00
44 TM	Trichloroethene	0.367	0.378	-3.0	71	0.00
45 C	1,2-Dichloropropane	0.336	0.361	-7.4#	75	0.00
46 T	Dibromomethane	0.186	0.194	-4.3	75	0.00
47 T	Bromodichloromethane	0.465	0.506	-8.8	75	0.00
48 T	Methyl methacrylate	0.243	0.252	-3.7	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
 Data File : VY005686.D
 Acq On : 16 Aug 2021 22:07
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 05:22:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	77	0.00
50 S	Toluene-d8	1.174	1.231	-4.9	69	0.00
51 T	4-Methyl-2-Pentanone	0.283	0.291	-2.8	79	0.00
52 CM	Toluene	0.856	0.904	-5.6#	72	0.00
53 T	t-1,3-Dichloropropene	0.520	0.542	-4.2	73	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.595	-5.5	72	0.00
55 T	1,1,2-Trichloroethane	0.269	0.282	-4.8	74	0.00
56 T	Ethyl methacrylate	0.402	0.419	-4.2	76	0.00
57 T	1,3-Dichloropropane	0.474	0.498	-5.1	75	0.00
58 T	2-Chloroethyl Vinyl ether	0.192	0.205	-6.8	75	0.00
59 T	2-Hexanone	0.200	0.203	-1.5	77	0.00
60 T	Dibromochloromethane	0.333	0.352	-5.7	75	0.00
61 T	1,2-Dibromoethane	0.255	0.263	-3.1	74	0.00
62 S	4-Bromofluorobenzene	0.389	0.419	-7.7	72	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	70	0.00
64 T	Tetrachloroethene	0.381	0.401	-5.2	76	0.00
65 PM	Chlorobenzene	1.006	1.056	-5.0	74	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.404	-6.6	75	0.00
67 C	Ethyl Benzene	1.824	1.927	-5.6#	73	0.00
68 T	m/p-Xylenes	0.709	0.747	-5.4	73	0.00
69 T	o-Xylene	0.681	0.720	-5.7	74	0.00
70 T	Styrene	1.146	1.228	-7.2	75	0.00
71 P	Bromoform	0.246	0.250	-1.6	74	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.490	3.599	-3.1	73	0.00
74 T	N-amyl acetate	1.051	1.073	-2.1	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.671	0.663	1.2	77	0.00
76 T	1,2,3-Trichloropropane	0.518	0.443	14.5	67	0.00
77 T	Bromobenzene	0.857	0.887	-3.5	75	0.00
78 T	n-propylbenzene	4.145	4.326	-4.4	74	0.00
79 T	2-Chlorotoluene	2.328	2.440	-4.8	75	0.00
80 T	1,3,5-Trimethylbenzene	2.896	3.026	-4.5	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.257	0.239	7.0	74	0.00
82 T	4-Chlorotoluene	2.408	2.556	-6.1	76	0.00
83 T	tert-Butylbenzene	2.599	2.673	-2.8	73	0.00
84 T	1,2,4-Trimethylbenzene	2.858	3.046	-6.6	76	0.00
85 T	sec-Butylbenzene	3.749	3.892	-3.8	73	0.00
86 T	p-Isopropyltoluene	3.219	3.349	-4.0	73	0.00
87 T	1,3-Dichlorobenzene	1.673	1.765	-5.5	76	0.00
88 T	1,4-Dichlorobenzene	1.660	1.744	-5.1	77	0.00
89 T	n-Butylbenzene	2.916	3.048	-4.5	74	0.00
90 T	Hexachloroethane	0.579	0.618	-6.7	76	0.00
91 T	1,2-Dichlorobenzene	1.478	1.566	-6.0	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.123	3.1	81	0.00
93 T	1,2,4-Trichlorobenzene	1.015	1.054	-3.8	78	0.00
94 T	Hexachlorobutadiene	0.657	0.690	-5.0	75	0.00
95 T	Naphthalene	2.001	2.043	-2.1	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081621\
Data File : VY005686.D
Acq On : 16 Aug 2021 22:07
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 17 05:22:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.949	-5.6	80	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6