

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
 Data File : VY010043.D
 Acq On : 16 Aug 2022 14:53
 Operator : KP/MD
 Sample : N4179-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-7211930

Quant Time: Aug 16 16:54:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 07:01:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	75313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	120642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	116292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	58124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	46918	62.306	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.620%	
35) Dibromofluoromethane	7.710	113	41467	53.002	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.000%	
50) Toluene-d8	10.179	98	135756	46.901	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.800%	
62) 4-Bromofluorobenzene	12.477	95	48006	47.219	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.440%	
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	7961	7.827	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
Data File : VY010043.D
Acq On : 16 Aug 2022 14:53
Operator : KP/MD
Sample : N4179-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-7211930

Quant Time: Aug 16 16:54:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 07:01:08 2022
Response via : Initial Calibration

