

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
 Data File : VY010046.D
 Acq On : 16 Aug 2022 16:04
 Operator : KP/MD
 Sample : N4197-04
 Misc : 5.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-203-IDWSO-20220811-3

Quant Time: Aug 16 16:55:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 07:01:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	79320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	128835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	116957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	54532	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	42487	53.571	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.140%
35) Dibromofluoromethane	7.709	113	41373	49.519	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.040%
50) Toluene-d8	10.178	98	146326	47.338	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.483	95	44481	40.970	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
Data File : VY010046.D
Acq On : 16 Aug 2022 16:04
Operator : KP/MD
Sample : N4197-04
Misc : 5.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-203-IDWSO-20220811-3

Quant Time: Aug 16 16:55:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 07:01:08 2022
Response via : Initial Calibration

