

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
 Data File : VY010047.D
 Acq On : 16 Aug 2022 16:28
 Operator : KP/MD
 Sample : N4197-05
 Misc : 5.98g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-203-IDWSO-20220811-4

Quant Time: Aug 16 16:56:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 07:01:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	75762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	117433	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	104259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	44812	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	32276	42.608	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.220%
35) Dibromofluoromethane	7.710	113	35587	46.730	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.460%
50) Toluene-d8	10.173	98	133474	47.372	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.477	95	38609	39.014	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.020%

Target Compounds			Qvalue			
16) Acetone	3.936	43	2815	17.350	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
Data File : VY010047.D
Acq On : 16 Aug 2022 16:28
Operator : KP/MD
Sample : N4197-05
Misc : 5.98g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-203-IDWSO-20220811-4

Quant Time: Aug 16 16:56:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 07:01:08 2022
Response via : Initial Calibration

