

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
 Data File : VY010052.D
 Acq On : 16 Aug 2022 18:44
 Operator : KP/MD
 Sample : N4208-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 W-2

Quant Time: Aug 17 05:14:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	69840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	115727	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	116212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	56574	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	43669	62.536	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.080%
35) Dibromofluoromethane	7.710	113	39280	52.339	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.680%
50) Toluene-d8	10.173	98	131017	47.186	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.477	95	44758	45.894	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081622\
Data File : VY010052.D
Acq On : 16 Aug 2022 18:44
Operator : KP/MD
Sample : N4208-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
W-2

Quant Time: Aug 17 05:14:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

