

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081623\
 Data File : VY015153.D
 Acq On : 16 Aug 2023 16:29
 Operator : SY/MD
 Sample : 04022-01
 Misc : 4.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-16

Quant Time: Aug 17 05:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	126721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	223522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	221096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	109743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	84912	55.820	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.640%
35) Dibromofluoromethane	7.716	113	71783	49.919	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.840%
50) Toluene-d8	10.179	98	265935	49.679	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.477	95	99819	51.054	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.100%
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	14856	8.919	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081623\
Data File : VY015153.D
Acq On : 16 Aug 2023 16:29
Operator : SY/MD
Sample : 04022-01
Misc : 4.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-16

Quant Time: Aug 17 05:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

