

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081623\
 Data File : VY015159.D
 Acq On : 16 Aug 2023 18:50
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 05:33:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.253	0.205	19.0	75	0.00
3 P	Chloromethane	0.347	0.310	10.7	80	0.00
4 C	Vinyl Chloride	0.398	0.379	4.8#	79	0.00
5 T	Bromomethane	0.291	0.323	-11.0	87	0.00
6 T	Chloroethane	0.289	0.298	-3.1	82	0.00
7 T	Trichlorofluoromethane	0.670	0.651	2.8	80	0.00
8 T	Diethyl Ether	0.263	0.266	-1.1	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.417	0.409	1.9	81	0.00
10 T	Methyl Iodide	0.409	0.348	14.9	63	0.00
11 T	Tert butyl alcohol	0.116	0.099	14.7	80	0.01
12 CM	1,1-Dichloroethene	0.348	0.335	3.7#	77	0.00
13 T	Acrolein	0.056	0.030	46.4#	66	0.00
14 T	Allyl chloride	0.617	0.605	1.9	78	0.00
15 T	Acrylonitrile	0.178	0.190	-6.7	84	0.00
16 T	Acetone	0.196	0.181	7.7	70	0.00
17 T	Carbon Disulfide	0.612	0.578	5.6	73	0.00
18 T	Methyl Acetate	0.719	0.836	-16.3	92	0.00
19 T	Methyl tert-butyl Ether	1.479	1.484	-0.3	80	0.00
20 T	Methylene Chloride	0.570	0.530	7.0	87	0.00
21 T	trans-1,2-Dichloroethene	0.392	0.391	0.3	80	0.00
22 T	Diisopropyl ether	1.539	1.544	-0.3	80	0.00
23 T	Vinyl Acetate	0.918	0.869	5.3	74	0.00
24 P	1,1-Dichloroethane	0.833	0.851	-2.2	82	0.00
25 T	2-Butanone	0.278	0.280	-0.7	80	0.00
26 T	2,2-Dichloropropane	0.803	0.683	14.9	71	0.00
27 T	cis-1,2-Dichloroethene	0.536	0.545	-1.7	82	0.00
28 T	Bromochloromethane	0.383	0.365	4.7	77	0.00
29 T	Tetrahydrofuran	0.166	0.173	-4.2	83	0.00
30 C	Chloroform	0.934	0.963	-3.1#	83	0.00
31 T	Cyclohexane	0.597	0.540	9.5	77	0.00
32 T	1,1,1-Trichloroethane	0.795	0.792	0.4	80	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.652	-8.7	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.322	0.339	-5.3	83	0.00
36 T	1,1-Dichloropropene	0.356	0.345	3.1	80	0.00
37 T	Ethyl Acetate	0.332	0.325	2.1	82	0.00
38 T	Carbon Tetrachloride	0.421	0.404	4.0	79	0.00
39 T	Methylcyclohexane	0.399	0.375	6.0	77	0.00
40 TM	Benzene	1.076	1.065	1.0	81	0.00
41 T	Methacrylonitrile	0.171	0.177	-3.5	80	0.00
42 TM	1,2-Dichloroethane	0.391	0.392	-0.3	82	0.00
43 T	Isopropyl Acetate	0.604	0.585	3.1	79	0.00
44 TM	Trichloroethene	0.310	0.305	1.6	81	0.00
45 C	1,2-Dichloropropane	0.303	0.303	0.0	83	0.00
46 T	Dibromomethane	0.191	0.194	-1.6	82	0.00
47 T	Bromodichloromethane	0.457	0.459	-0.4	82	0.00
48 T	Methyl methacrylate	0.258	0.266	-3.1	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081623\
 Data File : VY015159.D
 Acq On : 16 Aug 2023 18:50
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 05:33:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	83	0.00
50 S	Toluene-d8	1.197	1.244	-3.9	83	0.00
51 T	4-Methyl-2-Pentanone	0.350	0.362	-3.4	83	0.00
52 CM	Toluene	0.709	0.699	1.4#	81	0.00
53 T	t-1,3-Dichloropropene	0.480	0.451	6.0	76	0.00
54 T	cis-1,3-Dichloropropene	0.505	0.481	4.8	78	0.00
55 T	1,1,2-Trichloroethane	0.291	0.293	-0.7	83	0.00
56 T	Ethyl methacrylate	0.426	0.421	1.2	79	0.00
57 T	1,3-Dichloropropane	0.475	0.472	0.6	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.198	0.195	1.5	82	0.00
59 T	2-Hexanone	0.259	0.263	-1.5	81	0.00
60 T	Dibromochloromethane	0.347	0.352	-1.4	82	0.00
61 T	1,2-Dibromoethane	0.276	0.271	1.8	80	0.00
62 S	4-Bromofluorobenzene	0.437	0.441	-0.9	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.348	0.380	-9.2	90	0.00
65 PM	Chlorobenzene	0.911	0.887	2.6	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.374	-1.1	83	0.00
67 C	Ethyl Benzene	1.596	1.538	3.6#	80	0.00
68 T	m/p-Xylenes	0.606	0.586	3.3	80	0.00
69 T	o-Xylene	0.604	0.584	3.3	80	0.00
70 T	Styrene	1.045	1.034	1.1	81	0.00
71 P	Bromoform	0.268	0.265	1.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.215	3.013	6.3	81	0.00
74 T	N-amyl acetate	1.161	1.088	6.3	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.837	0.810	3.2	81	0.00
76 T	1,2,3-Trichloropropane	0.607	0.609	-0.3	87	0.00
77 T	Bromobenzene	0.784	0.751	4.2	83	0.00
78 T	n-propylbenzene	3.818	3.606	5.6	82	0.00
79 T	2-Chlorotoluene	2.203	2.076	5.8	81	0.00
80 T	1,3,5-Trimethylbenzene	2.680	2.531	5.6	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.277	0.246	11.2	73	0.00
82 T	4-Chlorotoluene	2.275	2.178	4.3	82	0.00
83 T	tert-Butylbenzene	2.451	2.304	6.0	80	0.00
84 T	1,2,4-Trimethylbenzene	2.678	2.563	4.3	81	0.00
85 T	sec-Butylbenzene	3.592	3.430	4.5	82	0.00
86 T	p-Isopropyltoluene	2.978	2.867	3.7	82	0.00
87 T	1,3-Dichlorobenzene	1.554	1.495	3.8	82	0.00
88 T	1,4-Dichlorobenzene	1.559	1.513	3.0	83	0.00
89 T	n-Butylbenzene	2.808	2.658	5.3	80	0.00
90 T	Hexachloroethane	0.591	0.555	6.1	81	0.00
91 T	1,2-Dichlorobenzene	1.452	1.431	1.4	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.162	0.159	1.9	83	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.873	6.3	80	0.00
94 T	Hexachlorobutadiene	0.507	0.469	7.5	81	0.00
95 T	Naphthalene	2.322	2.203	5.1	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081623\
Data File : VY015159.D
Acq On : 16 Aug 2023 18:50
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 17 05:33:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.842	0.810	3.8	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5