

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
 Data File : VY019086.D
 Acq On : 16 Aug 2024 12:08
 Operator : SY/MD
 Sample : P3607-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOILB
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-08142024RE

Quant Time: Aug 16 14:27:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	14890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	22484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	14699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	3684	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	4893	32.782	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	65.560%
35) Dibromofluoromethane	7.622	113	6681	45.445	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.880%
50) Toluene-d8	10.103	98	22540	42.605	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.200%
62) 4-Bromofluorobenzene	12.408	95	4727	26.210	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	52.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
Data File : VY019086.D
Acq On : 16 Aug 2024 12:08
Operator : SY/MD
Sample : P3607-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOILB
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-08142024RE

Quant Time: Aug 16 14:27:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

