

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
 Data File : VY019093.D
 Acq On : 16 Aug 2024 15:01
 Operator : SY/MD
 Sample : P3615-07
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-04-081424

Quant Time: Aug 16 23:14:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.70
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.60
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.41
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	58 - 134	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
Data File : VY019093.D
Acq On : 16 Aug 2024 15:01
Operator : SY/MD
Sample : P3615-07
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-04-081424

Quant Time: Aug 16 23:14:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

