

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
 Data File : VY019094.D
 Acq On : 16 Aug 2024 15:26
 Operator : SY/MD
 Sample : P3580-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IBLK

Quant Time: Aug 16 23:14:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.689	168	921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	2001	50.000	ug/l	# 0.00
63) Chlorobenzene-d5	11.413	117	1592	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	208	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.042	65	1003	108.640	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	217.280%#
35) Dibromofluoromethane	7.628	113	778	59.463	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.920%
50) Toluene-d8	10.097	98	2304	48.934	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.401	95	253	15.762	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	31.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
Data File : VY019094.D
Acq On : 16 Aug 2024 15:26
Operator : SY/MD
Sample : P3580-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IBLK

Quant Time: Aug 16 23:14:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

