

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
 Data File : VY019095.D
 Acq On : 16 Aug 2024 15:51
 Operator : SY/MD
 Sample : P3580-02
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Aug 16 23:15:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	1896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	3197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	2171	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	1341	70.557	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 141.120%	
35) Dibromofluoromethane	7.628	113	1031	49.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.640%	
50) Toluene-d8	10.109	98	3097	41.169	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 82.340%	
62) 4-Bromofluorobenzene	12.402	95	577	22.500	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 45.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081624\
Data File : VY019095.D
Acq On : 16 Aug 2024 15:51
Operator : SY/MD
Sample : P3580-02
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Aug 16 23:15:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

