

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081721\
 Data File : VY005721.D
 Acq On : 17 Aug 2021 19:25
 Operator : SY/MD
 Sample : M3430-03
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-3

Quant Time: Aug 18 01:53:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	33495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	51269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	49406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	23016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	19625	57.882	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 115.760%	
35) Dibromofluoromethane	7.728	113	15500	53.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.000%	
50) Toluene-d8	10.185	98	64378	53.477	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.960%	
62) 4-Bromofluorobenzene	12.483	95	20130	50.495	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.000%	
Target Compounds						
20) Methylene Chloride	4.722	84	6616	15.320	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081721\
Data File : VY005721.D
Acq On : 17 Aug 2021 19:25
Operator : SY/MD
Sample : M3430-03
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-3

Quant Time: Aug 18 01:53:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

