

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010061.D
 Acq On : 17 Aug 2022 12:20
 Operator : KP/MD
 Sample : N4218-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-081522

Quant Time: Aug 17 14:57:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	65584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	105041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	114437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	55268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	39787	60.674	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.340%
35) Dibromofluoromethane	7.710	113	36053	52.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.860%
50) Toluene-d8	10.179	98	133375	52.922	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.840%
62) 4-Bromofluorobenzene	12.483	95	46515	52.548	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.100%
Target Compounds						
16) Acetone	3.948	43	4364	31.070	ug/l #	88
20) Methylene Chloride	4.710	84	5462	5.400	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010061.D
Acq On : 17 Aug 2022 12:20
Operator : KP/MD
Sample : N4218-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-081522

Quant Time: Aug 17 14:57:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

