

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010064.D
 Acq On : 17 Aug 2022 13:30
 Operator : KP/MD
 Sample : N4143-05
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-26414

Quant Time: Aug 17 14:57:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	69477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	110210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	119946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	60354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40979	58.990	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.980%
35) Dibromofluoromethane	7.709	113	37784	52.866	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.740%
50) Toluene-d8	10.178	98	136245	51.525	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.483	95	51307	55.243	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.480%
Target Compounds						
16) Acetone	3.936	43	1277	8.582	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010064.D
Acq On : 17 Aug 2022 13:30
Operator : KP/MD
Sample : N4143-05
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-26414

Quant Time: Aug 17 14:57:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

