

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010066.D
 Acq On : 17 Aug 2022 14:18
 Operator : KP/MD
 Sample : N4191-04
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-367

Quant Time: Aug 18 08:19:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	73087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	115688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	120979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40676	55.662	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.320%
35) Dibromofluoromethane	7.710	113	39958	53.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.520%
50) Toluene-d8	10.179	98	134801	48.565	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.140%
62) 4-Bromofluorobenzene	12.477	95	52677	54.032	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.060%
Target Compounds						
						Qvalue
16) Acetone	3.936	43	3339	21.332	ug/l	# 89
68) m/p-Xylenes	11.697	106	2392	1.424	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010066.D
Acq On : 17 Aug 2022 14:18
Operator : KP/MD
Sample : N4191-04
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-367

Quant Time: Aug 18 08:19:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

