

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010071.D
 Acq On : 17 Aug 2022 16:15
 Operator : KP/MD
 Sample : N3894-13
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-32

Quant Time: Aug 18 08:20:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	69674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	116045	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	116325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	55714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	42511	61.022	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.040%
35) Dibromofluoromethane	7.710	113	39531	52.529	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.060%
50) Toluene-d8	10.179	98	131401	47.195	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.477	95	46147	47.189	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.380%
Target Compounds						
						Qvalue
16) Acetone	3.936	43	1847	12.378	ug/l	# 88
20) Methylene Chloride	4.704	84	6520	6.514	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010071.D
Acq On : 17 Aug 2022 16:15
Operator : KP/MD
Sample : N3894-13
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-32

Quant Time: Aug 18 08:20:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

