

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010076.D
 Acq On : 17 Aug 2022 18:12
 Operator : KP/MD
 Sample : N4217-10
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP4-0815

Quant Time: Aug 18 08:22:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	67213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	105572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	111780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	52681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	37902	56.398	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.800%
35) Dibromofluoromethane	7.710	113	38109	55.663	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.320%
50) Toluene-d8	10.179	98	121691	48.043	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.080%
62) 4-Bromofluorobenzene	12.477	95	42315	47.563	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.120%
Target Compounds						
16) Acetone	3.942	43	2595	18.028	ug/l	98
20) Methylene Chloride	4.710	84	7395	8.294	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010076.D
Acq On : 17 Aug 2022 18:12
Operator : KP/MD
Sample : N4217-10
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP4-0815

Quant Time: Aug 18 08:22:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

